

Government of West Bengal
Office of the Chief Medical Officer of Health, Jalpaiguri
District Health Administrative Building, 1st Floor, Hospital Road
Jalpaiguri-735101 Fax-03561-232001: Ph-03561-225394
E-mail- e-mail: cmohojal2020@gmail.com : cmohofficejalpaiguri@gmail.com

Memo No. 1741

Dated. 10/05/2022

Notice Inviting Quotation

Sealed quotations are hereby invited from the reputed bonafide Firms /Agencies/ Contractors by the undersigned for supplying the following article to the Chief Medical Officer of Health Office, Jalpaiguri. The rate should be quoted inclusive of all taxes and delivery charges. All quotationers are requested to submit their up to date papers (Trade License for 2022-2023, Pan no, GST Registration, P.Tax Certificate) along with the rate sheet. The authority reserve the right to cancel or reject the whole quotation or partly without assigning any reason thereof. The Quotation will be valid for the financial year 2022-2023.

The quotation should reach to this office by registered post or by hand on or before 19/05/2022 up to 2:00 pm. Opening date of quotation is on 19/05/2022 at 4:00 pm.

Sl No.	Particulars	Quantity	Quoted rate
1	<p>Carbon Monoxide Monitor Technical Essential Specifications:</p> <ol style="list-style-type: none"> 1. It should have color touch-screen display. 2. Response time should be <30 seconds to 90% FSD. 3. Recording and Intepreting result should be quick and easy. 4. It should have automatic calibration to ensure accurate results. 5. There should be provision of mouthpieces for excellent and low-cost infection control. 6. It should have storage of upto100 reading and personal profiles. 7. There should be familiar green, amber and red traffic light system for making CO levels instantly identifiable to patients. 8. It should have electrochemical sensor with $\pm 5\%$ repeatability and accuracy. 9. Sensor operating life should be 5 years (2 year warranty) 10.Sensor sensitivity should be 1ppm. 11.Its weigh should be in <250 gm 12.H2 cross sensitivity: < 12% <p>Desirable Specifications:</p> <ol style="list-style-type: none"> 1. It should be able to provide instant result in exact ppm for %COHB and %FCOHB. 2. There should be provision for adults, adolescents and pregnant woman testing mode. <p>Environmental Factor:</p> <ol style="list-style-type: none"> 1. Operation temperature range should be 0-40 degree Celsius. 2. Storage/transport temperature: 0-50° C 3. Operating/Storage/transport presuure : Atmospheric $\pm 10\%$ 4. Operating humidity : 15-90% non-condensing 5. Storage/transport humidity : 0-95% <p>Documentation:</p> <ul style="list-style-type: none"> ▪ User/ Technical/ Maintenance manual to be supplied in English 	01 No.	Rate to be quoted in the Letter head of the Agencie's

10/5/2022
Chief Medical Officer of Health
Jalpaiguri

Memo No. Quotation/1741 (7)

Dated, Jalpaiguri the 10/05/2022

Copy forwarded for information & request to kindly arrange for display this notice to their office Notice Board.

1. The Sabhadhipati, Zilla Parishad, Jalpaiguri
2. The District Magistrate, Jalpaiguri.
3. The Station Master, Jalpaiguri Railway Station, Jalpaiguri.
4. The Post Master, Head Post Office, Jalpaiguri
5. The Superintendent, District Hospital, Jalpaiguri.
6. The DSM, with the direction to publish the matter in official website of CMOH, Jal.
7. Office Notice Board.


**Chief Medical Officer of Health
Jalpaiguri**

